


# PHscan30 Pocket pH Tester

## USER MANUAL




## Overview

Thank you for selecting the PHscan series pocket pH tester. This user manual provides a step-by-step guide to help you operate the tester, please carefully read the following instructions before use.

### Installing the Batteries


1. Twist the electrode collar counter clockwise, pull the electrode away from the tester.






2. Insert two AAA alkaline batteries into the battery compartment, note polarity.




3. Push the electrode into the tester and twist the electrode collar clockwise until tight.



### Keypad

Key	Function
	<ul style="list-style-type: none"> <li>• Switch the tester on or off</li> <li>• Lock or unlock measurement</li> <li>• Exit the calibration, settings and return to the pH measurement</li> </ul>
	<ul style="list-style-type: none"> <li>• Start calibration</li> <li>• Press and hold the key to enter the setup menu</li> <li>• Select an option</li> </ul>
	<ul style="list-style-type: none"> <li>• Confirm the calibration, settings or displayed option</li> </ul>


### Display

Icon	Description
	When the battery voltage falls below the minimum power requirements, the icon automatically disappears

MEAS	Indicates that the tester is in the measurement mode
CAL	Indicates that the tester is in the calibration mode
SETUP	Indicates that the tester is in the setup mode
ATC	Indicates that the automatic temperature compensation is enabled

## Prior to Use

Remove the protective cap (and translucent cover) from the bottom of the tester. If some salt crystals deposited on the electrode, rinse with tap water to clean these deposits.




If tiny air bubbles are present inside the pH-sensitive glass membrane, gently shake the tester downward to remove air bubbles.



If the glass membrane has dried out, soak the electrode in 3M KCl or pH 4.01 buffer solution for about 30 minutes. If above solutions are not available, use the tap water.

## Switching the Tester On and Off

- Press and hold the **Meas** key for about 5 seconds to switch on the tester.
- Press and hold the **Meas** key to switch off the tester.



If you do not press any key within 8 minutes, the tester will switch off automatically to conserve energy.


## Setup Menu

The PHscan30 tester contains 7 menu items in the setup menu, the following table describes the functions of each option.

Menu Item	Option and Description
bUF	<b>pH Buffer Group</b> Set the pH buffer group for calibration and auto-recognition.
	USA USA (default)
	NIST NIST
CRL	<b>Calibration Points</b> Set the number of calibration points.
	1 1 point
	2 2 points (default)
	3 3 points
UNIT	<b>Measurement Unit</b> Set the default temperature unit.
	°C Degrees Celsius (default)
	°F Degrees Fahrenheit
CRL	<b>Temperature Calibration</b> Refer to page 4.
	°C Reading $\pm 10^{\circ}\text{C}$
	°F Reading $\pm 10^{\circ}\text{F}$
HOLD	<b>Auto-Hold</b> If enabled, the tester will automatically sense and lock the measurement endpoint.
	YES Enable
	NO Disable (default)
OFF	<b>Auto-Power Off</b> If enabled, the tester will automatically switch off if no key is pressed within 8 minutes.
	YES Enable (default)
	NO Disable
rSt	<b>Factory Reset</b> If enabled, all of the calibration values and current settings will be deleted or reset to the factory defaults, the tester must be recalibrated.
	YES Enable
	NO Disable (default)

## Setting the Default Option


1. In the measurement mode, press and hold the **Cal** key to enter the setup menu.



2. If necessary, press the **Cal** key again to select an option.



3. Press the **Enter** key, the tester saves the current option and moves to the next menu item.



4. Repeat steps above until the tester returns to the measurement mode.



- During the setting process, if you do not need to calibrate the temperature, press the **Enter** key to skip the °C/CRL or °F/CRL option.
- To exit the setup menu, press the **Meas** key.

## pH Calibration

The PHscan30 tester allows 1 to 3 points calibration, we recommend that you perform at least 2 points calibration for accurate pH measurement, the tester will automatically recognize and calibrate to following standard buffer values.

USA standard buffers	pH 4.01, 7.00, 10.01
NIST standard buffers	pH 4.01, 6.86, 9.18

Single point calibration should only be carried out with the pH 7.00 or 6.86, otherwise calibration will not be accepted.

For better accuracy, we recommend calibrating the tester regularly. DO NOT reuse the buffer solutions after calibration, contaminants in solution will affect the calibration and eventually the accuracy of the measurement.


## Single Point Calibration

Ensure that you have selected 1 point calibration in the setup menu.

- 1.1 Press the **Cal** key, the tester shows pH 7.00/CAL 1.



- 1.2 Rinse the electrode with distilled water, place the electrode into the pH 7.00 buffer solution, stir the tester gently to create a homogeneous solution.




- 1.3 Press the **Enter** key, the tester begins the calibration. When the reading has stabilized, the display will show **End**. Calibration is completed.



## 2 Points Calibration

Ensure that you have selected 2 points calibration in the setup menu.


- 2.1 Repeat steps 1.1 through 1.3 above. When the first calibration point is completed, the display will show ---- / CAL 2, the tester prompts you to continue with second point calibration.



- 2.2 Rinse the electrode with distilled water, place the electrode into pH 4.01 or 10.01 buffer solution, stir the tester gently.



- 2.3 Press the **Enter** key, the tester automatically recognizes the buffer solution and begins calibration. When the reading has stabilized, the display automatically shows the electrode slope (e.g., 99%) and **End**. Second point calibration is completed.



## 3 Points Calibration

Ensure that you have selected 3 points calibration in the setup menu.

- 3.1 Repeat steps 1.1 through 1.3 above. When the first calibration point is completed, the display will show pH 4.01/CAL 2, the tester prompts you to continue with second point calibration.
- 3.2 Rinse the electrode with distilled water and place into the pH 4.01 buffer solution, stir the tester gently. Press the **Enter** key, the tester begins calibration. When the reading has stabilized, the display will automatically show the electrode slope and pH 10.01/CAL 3.
- 3.3 Rinse the electrode with distilled water and place into the pH 10.01 buffer solution, stir the tester gently. Press the **Enter** key. When the reading has stabilized, the display will automatically show the electrode slope and **End**. Calibration is completed.



- During the calibration, if the display shows **Err** indicating that the measured mV value for the current calibration point deviates by more than 60 mV (approximately  $\pm 1$  pH) from the theoretical value of the pH buffer. The calibration will not be accepted. Please check the electrode and make sure the buffer solutions are fresh and uncontaminated.
- To exit the calibration without saving calibrated values, press the **Meas** key.


## Temperature Calibration

The PHscan30 tester is installed with a built-in temperature sensor for automatic temperature compensation. During the measurement, if the measured temperature reading differs from that of an accurate thermometer, the tester needs to be calibrated.

1. Press and hold the **Cal** key to enter the setup menu.



2. Press the **Enter** key until the display shows °C/°F or °C/°F.



3. Press the **Cal** key, the tester enters the temperature calibration mode.




4. Place the electrode into a solution with a known accurate temperature and wait for the measurement to stabilize.



5. Press the **Cal** key to modify the temperature value.




6. Press the **Enter** key to save and press the **Meas** key to return to the measurement mode.



## Measurement

Rinse the electrode with distilled water, place the electrode into the sample solution and stir gently. Note, DO NOT completely immerse the tester in water. Wait for the measurement to stabilize and record the reading.



If your tester installed a spear tip electrode or flat surface electrode (refer to the **Optional Accessories > pH Electrodes** section) and the samples are soft solid or semi-solid, wet the sample with clean water, then slight pressure the electrode to take measure. Note that the pH-sensitive membrane must be fully touched with sample surface.



- During the measurement, never wipe the pH-sensitive membrane as this will cause static interference, blot dry with a lint-free tissue to remove waterdrops on electrode.
- If the display shows ---- indicating the measurement exceeds the range, remove the tester from the sample immediately.
- If the **HOLD** option is enabled in the setup menu, the tester will automatically lock the measurement endpoint and show HOLD icon. Press the **Meas** key to resume measuring.



## Electrode Maintenance and Replacement

### Cleaning the Electrode

- Since the pH-sensitive membrane is susceptible to contamination, make sure that rinse the electrode thoroughly with distilled water after use.
- If your sample contains the oil or grease, soak the electrode in mild detergent or electrode cleaning solution for at least 15 minutes, then rinse with distilled water.
- If you do not use the tester for a period longer than 1 month, store the electrode in 3M KCl solution or electrode storage solution.



DO NOT store the electrode in distilled or deionized water, which will deplete the hydration layer of the pH-sensitive membrane and render the electrode useless.

### Replacing the Electrode

If the tester fails to calibrate or gives fluctuating readings, you should consider replacing the electrode.

1. Twist the electrode collar counter clockwise, pull the electrode away from the tester.



2. Align the slot on the new electrode, gently push the electrode into the tester.



3. Twist the electrode collar clockwise until tight.



## Appendix


### Preparation of pH Buffer Solutions

The PHscan30 tester is packaged with the pH 4.01/7.00/10.01 buffer sachets required for calibration.

1. Half fill a 250 ml volumetric flask with distilled water and add the pH 7.00 buffer reagent.



2. Swirl the volumetric flask gently to dissolve the reagent and fill to the mark with distilled water.



3. Cap and upend the volumetric flask several times to mix solution.







- Preparation of pH 4.01 and 10.01 buffer solutions are the same as above.
- Prepared buffer solution should be stored in hermetically sealed glass container and avoid direct sunlight.

### Preparation of Electrode Storage Solution

Dissolve 24.6 grams of analytical grade KCl reagent in 100 ml distilled water. Add pH 4.01 standard buffer and adjust solution to pH 4.

## Optional Accessories

### pH Electrodes

	Order Code and Description
	<b>E-PHscan-ST-10K</b> <ul style="list-style-type: none"> <li>• Circular pH-sensitive membrane</li> <li>• For measuring the general water samples (non-viscous, non-corrosive liquids)</li> </ul>
	<b>E-PHscan-FT-10K</b> <ul style="list-style-type: none"> <li>• Flat surface pH-sensitive membrane</li> <li>• For measuring the surfaces of semi-solid or gel samples, such as cream, doughs, paper, textiles, paints, coatings, printing inks, etc.</li> </ul>
	<b>E-PHscan-LT-10K</b> <ul style="list-style-type: none"> <li>• Circular pH-sensitive membrane</li> <li>• Electrode dimensions: 75 × 12.5 (Dia.) mm</li> <li>• For measuring the samples in small containers, such as test tube</li> </ul>
	<b>E-PHscan-PT-10K</b> <ul style="list-style-type: none"> <li>• Spear tip pH-sensitive membrane</li> <li>• For piercing and measuring the semi-solid samples, such as soil, vegetables, fruits, meats, cheeses, etc.</li> </ul>

### Solutions

Order Code	Description
PHCS-USA	pH 4.01, 7.00, 10.01 buffer solutions, 480 ml
PHCS-ES	Electrode storage solution, 480 ml
PHCS-A	Removes acidic deposits, 480 ml
PHCS-B	Removes bacterial contaminants, 480 ml
PHCS-G	Removes oil and grease, 480 ml
PHCS-O	Removes organic contaminants, 480 ml
PHCS-P	Removes protein residues, 480 ml

## Tester Specifications

Model	PHscan30
<b>pH</b>	
Range	-1.00 to 15.00 pH
Resolution	0.01 pH
Accuracy	±0.01 pH
Calibration Point	1 to 3 points
pH Buffer Option	USA or NIST

Automatic Buffer Recognition	pH 4.01, 7.00, 10.01 or 4.01, 6.86, 9.18
Temperature Compensation	0 to 60°C (32 to 140°F), automatic
<b>Temperature</b>	
Range	0 to 60°C (32 to 140°F)
Resolution	0.1°C (0.1°F)
Accuracy	±1°C (±1.8°F)
Calibration Point	1 point
Calibration Range	Reading ±10°C/°F
<b>Other Specifications</b>	
Operating Temperature	0 to 50°C (32 to 122°F)
Storage Temperature	0 to 60°C (32 to 140°F)
Relative Humidity	< 80% (non-condensing)
IP Rating	IP54
Display	LCD, 21 × 21 mm (0.82 × 0.82 in.)
Power Requirements	2 × 1.5V AAA alkaline batteries
Auto-Off	8 minutes after last key pressed
Dimensions	185 (L) × 40 (Ø) mm (7.28 × 1.57 in.)
Weight	100g (3.5 oz.)

## Disposal

This product is required to comply with the European Union's Waste Electrical and Electronic Equipment (WEEE) Directive 2002/96/EC and may not be disposed of in domestic waste. Please dispose of product in accordance with local regulations at the collecting point specified for electrical and electronic equipment.



## Warranty

The warranty period for tester is one year from the date of shipment. Above warranty does not cover the electrode and pH buffer solutions.

Out of warranty products will be repaired on a charged basis.

The warranty on your tester shall not apply to defects resulting from:

- Improper or inadequate maintenance by customer
- Unauthorized modification or misuse
- Operation outside of the environment specifications of the products

For more information, please contact the supplier.



Office: 4715 Castlewood St., Sugar land, TX 77479, USA

Tel: (+1) 346-762-7358

E-mail: banteinstruments@yahoo.com

Factory: F3, Building 2, No.2185, Laifang Rd., Shanghai 201615, China

Tel: (+86) 21-6404-1598

E-mail: banteinstrument@hotmail.com



[www.bante-china.com](http://www.bante-china.com)



The information in this document is subject to change without notice.  
Copyright © Bante Instruments Inc, 2022. All rights reserved.